



## CERTIFICATO CE

Certificato n. 2010/MDD

### Dichiarazione di approvazione del sistema qualità

(Sistema completo di garanzia qualità)

Visto l'esito delle verifiche condotte in conformità all'Allegato II, con l'esclusione del punto 4, della direttiva 93/42/CEE e s.m.i., si dichiara che la ditta:

#### PHARMAPLAST SAE

23512 ALEXANDRIA - AMRIA FREE ZONE (EGY) - Egypt

mantiene nello stabilimento di:

23512 ALEXANDRIA - AMRIA FREE ZONE (EGY) - Egypt

un sistema qualità che assicura la conformità dei seguenti prodotti:

#### Cerotti per la cura del piede con acido salicilico

Modd. come da documento 'Allegato al Certificato CE no. 2010/MDD - Elenco dei Dispositivi' rev. 0 del 2021/01/11; tale allegato costituisce parte integrante e sostanziale del presente certificato.

ai requisiti essenziali della direttiva suddetta ad essi applicabili (in tutte le fasi dalla progettazione al controllo finale) ed è sottoposta alla sorveglianza prevista dal punto 5 dell'Allegato II. Per i dispositivi in classe III questo certificato è valido solamente con il relativo certificato di esame CE della progettazione di Allegato II.4.

Riferimento pratiche IMQ:

DM17-0016693-01; DM19-0038241-01; DM20-0047908-01; DM20-0059504-01.

**Questa Dichiarazione di approvazione è rilasciata dall'IMQ S.p.A. quale organismo notificato per la direttiva 93/42/CEE e s.m.i. Il numero identificativo dell'IMQ S.p.A. quale organismo notificato è: 0051.**

Emesso il: 2018-10-29  
Data aggiornamento: 2021-01-11  
Sostituisce: 2020-01-27  
Data scadenza: 2023-10-28

IMQ



## EC CERTIFICATE

Certificate No 2010/MDD

### Full Quality Assurance System Approval Certificate

On the basis of our examination carried out according to Annex II, excluding section 4, of the Directive 93/42/EEC and its revised version, we hereby certify that:

#### PHARMAPLAST SAE

23512 ALEXANDRIA - AMRIA FREE ZONE (EGY) - Egypt

manages in the factory of:

23512 ALEXANDRIA - AMRIA FREE ZONE (EGY) - Egypt

a quality assurance system ensuring the conformity of the following products:

#### Foot care plasters with salicylic acid

Type ref. as to document 'Annex of EC Certificate no. 2010/MDD - Device List' rev. 0 dated 2021/01/11; this annex is integral and substantial part of this certificate.

with the relevant essential requirements of the aforementioned directive (from design to final inspection and testing) and it is subject to surveillance as specified in section 5 of Annex II. For class III devices, this certificate is valid only with the relevant EC Design-Examination Certificate of Annex II.4.

Reference to IMQ files Nos:

DM17-0016693-01; DM19-0038241-01; DM20-0047908-01; DM20-0059504-01.

**This Approval Certificate is issued by IMQ S.p.A. as Notified Body for the Directive 93/42/EEC and its revised version. Notified Body notified to European Commission under number: 0051.**

Date: 2018-10-29  
 Updated: 2021-01-11  
 Substitution Date: 2020-01-27  
 Expiry Date: 2023-10-28

IMQ

# Allegato al Certificato CE n. 2010/MDD - Elenco dei Dispositivi

Annex of EC Certificate no. 2010/MDD - Device List

rev. 0 del/of 2021/01/11

<b>Categoria di dispositivo:</b> Device category:	<b>Cerotti per la cura del piede con acido salicilico</b> Foot-care plasters with salicylic acid
<b>Modello/i:</b> Model(s):	<b>CornXY</b> <b>CureWRXY</b> <b>CureWRDXY</b> <b>CornkitXY</b> <b>CorncollXY</b>
	where: X = it is product length [mm] and can assume the following values: from 1 to 1000 Y = it is product width [mm] and can assume the following values: from 1 to 1000
<b>Marca/Marche:</b> Trade marks:	<b>Callivoro</b> <b>Cure-Aid®</b> <b>Fusswohl</b> <b>Pasocare Specialist Plus</b> <b>Salvelox Med</b> <b>Salvequick Med</b> <b>Sensomed®</b> <b>Smartherapy</b>